

**Activation Date:** \_\_\_\_\_  
*(Date the pressure pad is put into use.)*

**Serial #:** \_\_\_\_\_

**Testing the System**  
 Test the falls alert system before each use and inspect the pressure pads and the monitor regularly to make sure they are not damaged. Test the pressure pads by pressing firmly on the pressure pad for 3 seconds. When you remove the pressure, the alarm will sound. Reset the alarm.

	Date	Shift	Initials	Comments
Daily Checklist for Pressure Pads				